NOTICE: This document contains references to Varian. Please note that Varian, Inc. is now part of Agilent Technologies. For more information, go to **www.agilent.com/chem.**

Agilent Technologies

Varian B.V.
Herculesweg 8
4330 EA
Middelburg

Micro-GC Modification Kit



Carriergas tube stop

North/South America
2700 Mitchell Drive
Walnut Creek
94598 California, USA
Tel: ++(1)9259392400
Fax: ++(1)9259452360 or
++(1)9259452344

Europe
P.O. Box 8033
4330 EA Middelburg
The Netherlands
Tel: ++(31)118671000
Fax: ++(31)118623193

Australia/East Asia
679 Springvale Road
Mulgrave, Victoria 3171
Australia
Tel: ++(61)395607133
Fax: ++(61)395607950

© 2009 Varian, Inc. All Rights Reserved

Printed in the Netherlands

CP501380 Rev: 02 December 2009

This kit CP740828 contains:

1) CP740029 Viton O-rings 2x
2) CP740209 Stop carrier tube
3) CP740210 Nut for stop carrier tube



1) Install the O-rings over the stop and nut.

2) Uninstall the GC-channel according the Channel exchange manual (CP501334).

3) Install the stop and nut on the carriergas tube.